

Title	Subject	Date	Page
	実測値		2
2014. 4. 11. SP-8 同中機 (社内) 3m			
9:15	chA (sample) D6042845 295.2 K	chB (2nd stage) DT-670 295.4 K	ホ-7 (太い) 4.6x10 Torr
10:15	8.350 K	7.420 K	
11:30	4.860	3.723	
11:50	4.777	3.653	
13:00	4.633	3.522	
2014. 4. 28. chA (sample) chB (2nd) ホ-7 6 4m			
9:50	D6042845	DT-670	
10:50	295.6 K	295.8 K	
11:20	8.442	7.638	
11:50	6.330	5.190	
13:00	5.877	4.808	
	5.769	4.730	
2014. 5. 1. 同中機			
set	chA	chB	P. T. P. W
10.00	100.00 K	8.24	3.6 30 0 29% W
50.00	50.00 K	48.06 K	= = = 17% W
100.0	100.00 K	88.24 K	= = = 21% W
300.0	300.01 K	177.82 K	= = = 38% W
400.0	400.01 K	200.72 K	= = = 41% W
495.0	495.01 K	240.24 K	= = = 49% W



NIKI GLASS CO., LTD.

仁木工芸株式会社

Name:

Title	Subject	Date	Page
		/ /	73
2014.4.30. SP-8 10ヶ分録(社内).			
CHA(P+100).		CHB(2nd to 2nd)	T-2 6M.
9:50	290.8K	290.0K,	4.8×10^{-5} Torr
10:20	206.2	182.0.	
10:50	47.2	45.6	
11:20	---	5.094	
11:50	---	4.808	
13:00	---	4.746	
5/7. 2nd 2nd 2nd 1:			
CHA(P+100)		CHB(2nd).	P.2.0 W.H.
Set	300.0K	300.01K	182.4K 3.6 25 0.6 35% W
	500.0K	500.02K	256.2K 2 2 2 50%
	600.0K	600.02K	290.6K 2 1 2 80%
	700.0K	700.02K	300.8K 2 2 2 95% W
	800.0K.	800.02K	878W
	1 hrs.		
 NIKI GLASS CO., LTD. 仁木工芸株式会社		Name: 	

Title	Subject	Date	Page
2014.5-8 SP-8 田中 隆 (和)	CH A (pt100) CH B (2nd stage)	/ /	3/3
9:10	298.0 K	298.2 K	高圧 ホース 6m 4.4 x 10 ⁻⁵ Torr
	4.70 K		
Set	2/3 3月 5日	P, 2, P U.	
300.0 K	CH A (pt500) 300.02	CH B (2nd) 170.6 K	3.6 24.06 35 ⁹ /100
500 K	500.02	250.4 K	2 2 2 88 ⁹ /100
700 K	700.02	270.8	2 2 2 73 ⁹ /100
800 K	800.01 1 4月 10日	310.6	2 2 2 87 ⁹ /100
5/9	CH A (pt100) CH B (2nd)		
9:30	288.7 K	288.2 K	
13:00	---	4.72 K	
Set	2/3 3月 5日	P, 2, P U.	
300 K	CH A (pt100) 300.02 K	CH B (2nd) 170.8 K	3.6.25 28 35 ⁹ /100
500 K	500.02 K	250.2 K	2 2 2 88 ⁹ /100
700 K	700.01 K	269.8 K	2 2 2 73 ⁹ /100
800 K	800.01	310.2	2 2 2 87 ⁹ /100
 NIKI GLASS CO., LTD. 仁木工芸株式会社	Name:		